



Certificate of Analysis


No. CA 19151

Product: Eichrom's Pb Resin Cartridges
Lot #: PBSR19D
Date of Manufacture: 6-Sep-2019

Eichrom's Pb Resin cartridge Lot PBSR19D has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, PBS073118, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.0 %
% Lead in Strip	> 99 %	100 %
Gravimetric Yield	80-120 %	96.7 %
Chemical Yield	80-120 %	92.5 %


Jill Bryant
Quality System Coordinator

September 9, 2019