

Certificate of Analysis

No. CA 19174

Product: Eichrom's 25mm Resolve™ Filters

Date of Manufacture: 9-OCT-2019

Lot #: RF100919

Eichrom's 25mm Resolve™ Filters Lot # RF100919 has been manufactured according to Eichrom standard operating procedures. The filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	38.6
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	35.3
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0



Jill Bryant
Quality System Coordinator

October 17, 2019