



**Certificate of Analysis**  
No. CA20133

Product: Eichrom's TRU Resin  
Lot #: TRA071520  
Date of Manufacture: 15-JUL-2020

Eichrom's TRU Resin Lot # TRA071520 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Eu in 3.1 N nitric acid	≥ 60	88.5
% <sup>241</sup> Am stripped in 4M HCl		
Least acceptable mean (4 rep)	95.0	99.2
Max. standard deviation	2.5	0.2
% <sup>241</sup> Am stripped in 0.1M ammonium bioxalate		
Maximum acceptable mean (4 rep)	1.0	0.3
Max. standard deviation	0.5	0.1
% <sup>239</sup> Pu stripped in 0.1M ammonium bioxalate		
Least acceptable mean (4 rep)	95.0	98.7
Max. standard deviation	2.5	0.6
% <sup>239</sup> Pu stripped in 4M HCl		
Maximum acceptable mean (4 rep)	1.0	0.1
Max. standard deviation	0.5	0.2

Jill Bryant  
Quality System Coordinator

July 23, 2020